

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.992 °C	28.1 °C	-1.892 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	33 %HR	3.882 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.220 %HR	47 %HR	-2.220 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.251 %HR	70 %HR	-4.251 %HR	± 5.0 %HR	± 1.9E+00 %HR