

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |             |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|-------------|
| 900.6 RPM                | 900 RPM          | -0.6 RPM             | ± 18 RPM                    | ±                                  | 7.6E-01 RPM |
| 999.4 RPM                | 1000 RPM         | 0.6 RPM              | ± 20 RPM                    | ±                                  | 1.2E+00 RPM |
| 1149.8 RPM               | 1150 RPM         | 0.2 RPM              | ± 23 RPM                    | ±                                  | 9.5E-01 RPM |
| 1499.5 RPM               | 1499 RPM         | -0.5 RPM             | ± 30 RPM                    | ±                                  | 1.8E+00 RPM |
| 1449.6 RPM               | 1450 RPM         | 0.4 RPM              | ± 29 RPM                    | ±                                  | 1.2E+00 RPM |
| 1599.2 RPM               | 1600 RPM         | 0.8 RPM              | ± 32 RPM                    | ±                                  | 2.0E+00 RPM |
| 1998.4 RPM               | 2000 RPM         | 1.6 RPM              | ± 40 RPM                    | ±                                  | 1.9E+00 RPM |
| 2298.0 RPM               | 2300 RPM         | 2.0 RPM              | ± 46 RPM                    | ±                                  | 1.8E+00 RPM |
| 2497.6 RPM               | 2500 RPM         | 2.4 RPM              | ± 50 RPM                    | ±                                  | 1.2E+00 RPM |
| 2596.2 RPM               | 2600 RPM         | 3.8 RPM              | ± 52 RPM                    | ±                                  | 2.1E+00 RPM |