

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
900.6 RPM	900 RPM	-0.6 RPM	± 18 RPM	±	7.6E-01 RPM
999.4 RPM	1000 RPM	0.6 RPM	± 20 RPM	±	1.2E+00 RPM
1149.8 RPM	1150 RPM	0.2 RPM	± 23 RPM	±	9.5E-01 RPM
1249.5 RPM	1250 RPM	0.5 RPM	± 30 RPM	±	1.8E+00 RPM
1449.6 RPM	1450 RPM	0.4 RPM	± 29 RPM	±	1.2E+00 RPM
1599.2 RPM	1600 RPM	0.8 RPM	± 32 RPM	±	2.0E+00 RPM
1998.4 RPM	2000 RPM	1.6 RPM	± 40 RPM	±	1.9E+00 RPM
2298.0 RPM	2300 RPM	2.0 RPM	± 46 RPM	±	1.8E+00 RPM
2497.6 RPM	2500 RPM	2.4 RPM	± 50 RPM	±	1.2E+00 RPM
2596.2 RPM	2600 RPM	3.8 RPM	± 52 RPM	±	2.1E+00 RPM