

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.140 °C        | -24.7 °C  | 0.440 °C      | ± 0.95 °C            | ± 4.0E-01 °C                |
| 0 °C              | -0.331 °C         | 0.0 °C    | 0.331 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 5 °C              | 4.933 °C          | 5.2 °C    | 0.267 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 10 °C             | 9.937 °C          | 10.1 °C   | 0.163 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 25 °C             | 24.938 °C         | 25.2 °C   | 0.262 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 30 °C             | 29.792 °C         | 29.7 °C   | -0.092 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |
| 40 °C             | 39.919 °C         | 39.8 °C   | -0.119 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |

