

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.140 °C	-24.6 °C	0.540 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.331 °C	-0.2 °C	0.131 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	4.933 °C	5.0 °C	0.067 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	9.937 °C	9.9 °C	-0.037 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	24.938 °C	25.1 °C	0.162 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.792 °C	29.6 °C	-0.192 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.919 °C	39.6 °C	-0.319 °C	± 0.45 °C	± 4.0E-01 °C

