

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
30 %HR	30.509 %HR	32.3 %HR	1.791 %HR	± 3 %HR	± 1.9E+00 %HR	
50 %HR	49.650 %HR	49.1 %HR	-0.550 %HR	± 3 %HR	± 1.9E+00 %HR	
70 %HR	70.223 %HR	68.0 %HR	-2.223 %HR	± 3 %HR	± 1.9E+00 %HR	
90 %HR	90.540 %HR	87.0 %HR	-3.540 %HR	± 5 %HR	± 1.9E+00 %HR	

