

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.882 °C	29.3 °C	-0.582 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.464 %HR	39.2 %HR	10.736 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.650 %HR	58.2 %HR	8.550 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.335 %HR	81.5 %HR	6.165 %HR	± 5.0 %HR	± 1.9E+00 %HR