

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.882 °C	29.8 °C	-0.082 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.464 %HR	31.0 %HR	2.536 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	49.650 %HR	50.8 %HR	1.150 %HR	± 2.0 %HR	± 1.9E+00 %HR
75 %HR	75.335 %HR	72.4 %HR	-2.935 %HR	± 2.0 %HR	± 1.9E+00 %HR