

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.852 °C	29.7 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.640 %HR	32.6 %HR	2.960 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.517 %HR	51.0 %HR	1.483 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.854 %HR	72.4 %HR	-2.454 %HR	± 3.0 %HR	± 1.9E+00 %HR