

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.852 °C	29.6 °C	-0.252 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.640 %HR	34 %HR	4.360 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.517 %HR	50 %HR	0.483 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.854 %HR	75 %HR	0.146 %HR	± 5.0 %HR	± 1.9E+00 %HR