

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.0839 µl	100 µl	-0.084 µl	± 0.6 µl	±	5.3E-02 µl
500.0185 µl	500.0 µl	-0.018 µl	± 0.6 µl	±	5.9E-02 µl
1,000.1372 µl	1000.0 µl	-0.137 µl	± 0.6 µl	±	7.8E-02 µl