

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
999.985 µl	1000 µl	0.015 µl	± 0.6 µl	±	5.9E-02 µl
2499.911 µl	2500 µl	0.089 µl	± 0.6 µl	±	1.1E-01 µl
5000.023 µl	5000 µl	-0.023 µl	± 0.6 µl	±	1.7E-01 µl