

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
29.00 mm	29.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
28.50 mm	28.51 mm	10 µm	± 30 µm	±	5.8E+00 µm
28.00 mm	28.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
27.50 mm	27.51 mm	10 µm	± 30 µm	±	5.8E+00 µm
27.00 mm	27.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
26.50 mm	26.51 mm	10 µm	± 30 µm	±	5.8E+00 µm
26.00 mm	26.01 mm	10 µm	± 30 µm	±	5.8E+00 µm
25.50 mm	25.51 mm	10 µm	± 30 µm	±	5.8E+00 µm
25.00 mm	25.01 mm	10 µm	± 30 µm	±	5.8E+00 µm