

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.2	0.19933	0.670	± 1.3	± 4.3E-01 mm					
0 a 0.3	0.29932	0.680	± 1.4	± 4.3E-01 mm					
0 a 0.4	0.39910	0.900	± 1.5	± 4.3E-01 mm					
0 a 0.5	0.49909	0.910	± 1.6	± 4.3E-01 mm					
0 a 0.6	0.59907	0.930	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.69912	0.880	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.79907	0.930	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89929	0.710	± 1.9	± 4.3E-01 mm					
0 a 1	0.99907	0.930	± 2	± 4.3E-01 mm					
0 a 2	1.99914	0.860	± 2.8	± 4.8E-01 mm					
0 a 3	2.99981	0.850	± 3.6	± 4.8E-01 mm					
0 a 4	4.00007	0.550	± 4.4	± 4.8E-01 mm					
0 a 6	6.00016	0.370	± 6	± 3.6E-01 mm					
0 a 10	10.00015	-0.154	± 9.2	± 3.6E-01 mm					