

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.20 μm	24 μm	0.8 μm	± 5 μm	±	5.8E-01 μm
51.10 μm	54 μm	2.9 μm	± 5 μm	±	5.8E-01 μm
125.10 μm	126 μm	0.90 μm	± 6 μm	±	5.8E-01 μm
249.40 μm	248 μm	-1.40 μm	± 12 μm	±	5.8E-01 μm