

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------|----------------------|--------------------------|------------------------------------|
| 23.2 μm | 25.7 μm | 2.5 μm | $\pm 2.5 \mu\text{m}$ | $\pm 5.8\text{E-}02 \mu\text{m}$ |
| 51.1 μm | 53.4 μm | 2.3 μm | $\pm 2.5 \mu\text{m}$ | $\pm 5.8\text{E-}02 \mu\text{m}$ |
| 125.1 μm | 127 μm | 1.9 μm | $\pm 2.5 \mu\text{m}$ | $\pm 6.1\text{E-}02 \mu\text{m}$ |
| 249.4 μm | 254 μm | 4.6 μm | $\pm 2.5 \mu\text{m}$ | $\pm 6.9\text{E-}02 \mu\text{m}$ |
| 478 μm | 474 μm | -4.0 μm | $\pm 2.5 \mu\text{m}$ | $\pm 9.3\text{E-}02 \mu\text{m}$ |
| 952 μm | 961 μm | 9.0 μm | $\pm 2.5 \mu\text{m}$ | $\pm 1.6\text{E-}01 \mu\text{m}$ |