

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	25.7 μm	2.5 μm	± 2.5 μm	±	5.8E-02 μm
51.1 μm	53.4 μm	2.3 μm	± 2.5 μm	±	5.8E-02 μm
125.1 μm	127 μm	1.9 μm	± 3.8 μm	±	6.1E-02 μm
249.4 μm	254 μm	4.6 μm	± 7.5 μm	±	6.9E-02 μm
478 μm	474 μm	-4.0 μm	± 14.3 μm	±	9.3E-02 μm
952 μm	961 μm	9.0 μm	± 28.6 μm	±	1.6E-01 μm