

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.002 °C	28.6 °C	-1.402 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.946 %HR	35 %HR	5.054 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.568 %HR	51 %HR	1.432 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.630 %HR	77 %HR	2.370 %HR	± 5.0 %HR	± 1.9E+00 %HR