

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.002 °C	30.2 °C	0.198 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.946 %HR	31 %HR	1.054 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.568 %HR	46 %HR	-3.568 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.630 %HR	72 %HR	-2.630 %HR	± 5.0 %HR	± 1.9E+00 %HR