

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.5 °C	-0.210 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.181 °C	-0.2 °C	-0.019 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.023 °C	4.9 °C	-0.123 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.0 °C	-0.007 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.3 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.2 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.521 °C	40.4 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C

