

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.6 °C	-0.310 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.181 °C	-0.3 °C	-0.119 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.023 °C	4.7 °C	-0.323 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.8 °C	-0.207 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.2 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.521 °C	40.3 °C	-0.221 °C	± 0.5 °C	± 4.0E-01 °C

