

Patrón (Standard)	UBP (UT)	Error (Error)		Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
3600.507 s	3601 s	0.493 s	0.0137%	± 0.0116%	±	5.6E-02 s
3600.152 s	3601 s	0.848 s	0.0236%	± 0.0116%	±	5.6E-02 s
3600.548 s	3601 s	0.452 s	0.0126%	± 0.0116%	±	5.6E-02 s