

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.982 °C	29.3 °C	-0.682 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.314 %HR	31 %HR	0.686 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.985 %HR	47 %HR	-1.985 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.814 %HR	71 %HR	-3.814 %HR	± 5.0 %HR	± 1.9E+00 %HR