

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1200.452 s	1200 s	-0.452 s	-0.0377%	± 0.0116%	±	5.1E-02 s
1200.143 s	1200 s	-0.143 s	-0.0119%	± 0.0116%	±	5.1E-02 s
1200.423 s	1200 s	-0.423 s	-0.0352%	± 0.0116%	±	5.1E-02 s