

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| # 25 °C                    | 24.828 °C                  | 25.1 °C            | 0.272 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| # 50 °C                    | 49.957 °C                  | 50.2 °C            | 0.243 °C               | ± 0.5 °C                      | ± 1.1E+00 °C                         |
| # 70 °C                    | 69.541 °C                  | 69.5 °C            | -0.041 °C              | ± 0.5 °C                      | ± 1.1E+00 °C                         |
| 30 °C                      | 29.651 °C                  | 30.1 °C            | 0.449 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.905 %HR                 | 32.1 %HR           | 2.195 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 49.435 %HR                 | 49.0 %HR           | -0.435 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.284 %HR                 | 68.0 %HR           | -2.284 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |