

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.0625 in	0.0000 in	± 0.0031 in	±	5.8E-02 in
0.1000 in	0.0997 in	-0.0003 in	± 0.005 in	±	5.8E-02 in
0.1250 in	0.1247 in	-0.0003 in	± 0.0063 in	±	5.8E-02 in
0.2000 in	0.1998 in	-0.0002 in	± 0.0100 in	±	5.8E-02 in
0.2500 in	0.2502 in	0.0002 in	± 0.0125 in	±	5.8E-02 in