

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	23 μm	-0.3 μm	± 5 μm	±	5.8E-01 μm
51.6 μm	52 μm	0.4 μm	± 5 μm	±	5.8E-01 μm
124.6 μm	123 μm	-1.6 μm	± 6.2 μm	±	5.8E-01 μm
250.3 μm	249 μm	-1.3 μm	± 13 μm	±	5.8E-01 μm
486.0 μm	485 μm	-1.0 μm	± 24 μm	±	5.8E-01 μm