

Certificado No. (Certificate #):		57913		Equipo (Instrument):		Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-06-16		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
0.0500 in		0.0492 in		-0.0008 in		± 0.0001 in		±	4.2E-05 in
0.1000 in		0.0992 in		-0.0008 in		± 0.0001 in		±	4.9E-05 in
0.1010 in		0.1002 in		-0.0008 in		± 0.0001 in		±	4.7E-05 in
0.1050 in		0.1043 in		-0.0007 in		± 0.0001 in		±	4.4E-05 in
0.1100 in		0.1092 in		-0.0008 in		± 0.0001 in		±	5.5E-05 in
0.1500 in		0.1491 in		-0.0009 in		± 0.0001 in		±	6.8E-05 in
0.2000 in		0.1991 in		-0.0009 in		± 0.0001 in		±	5.5E-05 in
0.3000 in		0.2991 in		-0.0009 in		± 0.0001 in		±	5.2E-05 in
0.4000 in		0.3991 in		-0.0009 in		± 0.0001 in		±	6.9E-05 in
0.5000 in		0.4992 in		-0.0008 in		± 0.0001 in		±	7.8E-05 in
1.0000 in		0.9990 in		-0.0010 in		± 0.0001 in		±	3.6E-05 in