

Certificado No. (Certificate #):		57914		Equipo (Instrument):		Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-06-16		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
0.0500 in		0.0500 in		0.0000 in		± 0.0001 in		±	3.8E-05 in
0.1000 in		0.1000 in		0.0000 in		± 0.0001 in		±	3.9E-05 in
0.1010 in		0.1010 in		0.0000 in		± 0.0001 in		±	3.7E-05 in
0.1050 in		0.1050 in		0.0000 in		± 0.0001 in		±	3.8E-05 in
0.1100 in		0.1100 in		0.0000 in		± 0.0001 in		±	3.8E-05 in
0.1500 in		0.1500 in		0.0000 in		± 0.0001 in		±	3.6E-05 in
0.2000 in		0.2000 in		0.0000 in		± 0.0001 in		±	2.9E-05 in
0.3000 in		0.3000 in		0.0000 in		± 0.0001 in		±	3.1E-05 in
0.4000 in		0.4000 in		0.0000 in		± 0.0001 in		±	3.6E-05 in
0.5000 in		0.5000 in		0.0000 in		± 0.0001 in		±	2.9E-05 in
1.0000 in		0.9999 in		-0.0001 in		± 0.0001 in		±	3.6E-05 in