

Certificado No. (Certificate #):57916Equipo (Instrument):Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):2025-06-17Marca (Brand):Mitutoyo

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	0.1000 in	0.1000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	0.5000 in	0.5000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	4.0000 in	4.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	6.0000 in	6.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	7.5000 in	7.5000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Interior / (Interior)	3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Interior / (Interior)	5.0000 in	5.0000 in	0.0000 in	± 0.0012 in	± 6.8E-04 in
Profundidad / (Depth)	1.0000 in	1.0000 in	0.0000 in	± 0.0012 in	± 6.7E-04 in
Profundidad / (Depth)	2.0000 in	2.0000 in	0.0000 in	± 0.0012 in	± 6.7E-04 in