

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.4 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.001 °C	15.4 °C	0.399 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.3 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.369 °C	32.6 °C	0.231 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.4 %HR	0.830 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.9 %HR	1.452 %HR	± 3.0 %HR	± 1.9E+00 %HR