

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.041 °C	-0.4 °C	-0.359 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.8 °C	-0.207 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.3 °C	0.265 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.2 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.521 °C	40.4 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C