

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.052 °C	29.2 °C	-0.852 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.633 %HR	23.6 %HR	-8.033 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	49.496 %HR	40.9 %HR	-8.596 %HR	± 2.5 %HR	± 1.9E+00 %HR
75 %HR	75.396 %HR	43.1 %HR	-32.296 %HR	± 2.5 %HR	± 1.9E+00 %HR