

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P	Incertidumbre <i>(Uncertainty)</i>	
250.0755 ml	250 ml	-0.075 ml	± 0.1 ml	±	8.0E+00 µl