

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.1081 ml	50.0 ml	-0.108 ml	± 0.5 ml	±	3.5E-01 µl
100.2426 ml	100.0 ml	-0.243 ml	± 1.0 ml	±	3.5E-01 µl
499.6676 ml	500.0 ml	0.332 ml	± 2.5 ml	±	3.5E-01 µl
1,000.278 ml	1000 ml	-0.278 ml	± 5.0 ml	±	3.5E-01 µl