

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.0961 ml	50.0 ml	-0.096 ml	± 10 ml	±	3.5E-01 µl
100.2420 ml	100.0 ml	-0.242 ml	± 10 ml	±	3.5E-01 µl
499.8882 ml	500.0 ml	0.112 ml	± 10 ml	±	3.5E-01 µl
1,000.077 ml	1000.0 ml	-0.077 ml	± 10 ml	±	3.5E-01 µl