

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.4254 µl	10 µl	-0.4254 µl	± 0.12 µl	±	1.2E+00 µl
49.7212 µl	50 µl	0.2788 µl	± 0.5 µl	±	1.2E+00 µl
99.4423 µl	100 µl	0.5577 µl	± 8 µl	±	2.7E-01 µl