

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(M.P.E)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|----------------------|------------------------------------|------------|
| 9.91416 µl               | 10 µl            | 0.08584 µl           | ± 0.12 µl            | ±                                  | 5.1E-02 µl |
| 50.46297 µl              | 50 µl            | -0.46297 µl          | ± 0.5 µl             | ±                                  | 5.2E-02 µl |
| 100.24428 µl             | 100 µl           | -0.24428 µl          | ± 8 µl               | ±                                  | 5.3E-02 µl |