

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
1.0269 ml	1 ml	-0.0269 ml	± 0.007 ml	±	2.0E+01 µl
4.9360 ml	5 ml	0.0640 ml	± 0.03 ml	±	3.0E-01 µl
9.9683 ml	10 ml	0.0317 ml	± 0.05 ml	±	5.9E-01 µl