

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00244 ml	1 ml	-0.00244 ml	± 0.008 ml	±	7.8E-02 µl
4.97963 ml	5 ml	0.02037 ml	± 0.04 ml	±	3.0E-01 µl
9.96829 ml	10 ml	0.03171 ml	± 0.06 ml	±	5.9E-01 µl