

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.7431 µl	100 µl	0.2569 µl	± 0.8 µl	±	4.5E-01 µl
499.6175 µl	500 µl	0.3825 µl	± 4 µl	±	3.3E-01 µl
1000.638 µl	1000 µl	-0.6384 µl	± 8 µl	±	2.4E+00 µl