

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.95357 µl	100 µl	0.04643 µl	± 0.8 µl	±	5.3E-02 µl
499.21652 µl	500 µl	0.78348 µl	± 4 µl	±	5.9E-02 µl
999.63596 µl	1000 µl	0.36404 µl	± 8 µl	±	7.8E-02 µl