

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.1000 ml	0.1 ml	-0.00002 ml	± 0.05 ml	±	5.1E-02 µl
1.0003 ml	1.0 ml	-0.00026 ml	± 0.05 ml	±	6.0E-02 µl
10.0242 ml	10.0 ml	-0.02416 ml	± 0.05 ml	±	3.2E-01 µl
25.0065 ml	25.0 ml	-0.00650 ml	± 0.05 ml	±	8.1E-01 µl