

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.1000 ml	0.1 ml	-0.00001 ml	± 1 ml	±	5.1E-02 µl
1.0007 ml	1.0 ml	-0.00071 ml	± 1 ml	±	5.9E-02 µl
10.0172 ml	10.0 ml	-0.01719 ml	± 1 ml	±	3.1E-01 µl
25.0046 ml	25.0 ml	-0.00464 ml	± 2 ml	±	7.7E-01 µl