

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.982 °C	30.0 °C	0.058 °C	± 0.2 °C	± 4.0E-01 °C
30 %HR	29.292 %HR	32.5 %HR	3.208 %HR	± 3.5 %HR	± 1.9E+00 %HR
50 %HR	48.985 %HR	45.6 %HR	-3.385 %HR	± 3.5 %HR	± 1.9E+00 %HR
75 %HR	74.405 %HR	71.2 %HR	-3.205 %HR	± 3.5 %HR	± 1.9E+00 %HR