

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.251 °C	0.3 °C	0.551 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.814 °C	20.2 °C	0.386 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.869 °C	40.0 °C	0.131 °C	± 0.5 °C	± 4.0E-01 °C