

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.273 °C	29.6 °C	-0.673 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.016 %HR	32.0 %HR	2.984 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.087 %HR	47.0 %HR	-2.087 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.028 %HR	72.0 %HR	-3.028 %HR	± 5.0 %HR	± 1.9E+00 %HR