

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.273 °C                  | 30.2 °C            | -0.073 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.016 %HR                 | 31.0 %HR           | 1.984 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 49.087 %HR                 | 47.0 %HR           | -2.087 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.028 %HR                 | 75.0 %HR           | -0.028 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |