

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.273 °C	29.5 °C	-0.773 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.016 %HR	32.0 %HR	2.984 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.087 %HR	46.0 %HR	-3.087 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.028 %HR	71.0 %HR	-4.028 %HR	± 5.0 %HR	± 1.9E+00 %HR