

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.273 °C	30.3 °C	0.027 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.016 %HR	29.8 %HR	0.784 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.087 %HR	45.3 %HR	-3.787 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.028 %HR	68.6 %HR	-6.428 %HR	± 5.0 %HR	± 1.9E+00 %HR